

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000264  
10:42 22/07/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:01  
PRESSURIZING 3:05  
STERILIZING 4:00  
136.0 C 220kPa 4:00  
136.1 C 221kPa 4:30  
136.3 C 222kPa 5:00  
136.2 C 221kPa 5:30  
136.3 C 221kPa 6:00  
136.3 C 223kPa 6:30  
136.2 C 222kPa 7:00  
136.4 C 224kPa 7:30  
136.4 C 224kPa 7:30  
VENTING 7:30  
AIR DRYING 8:19  
CYCLE COMPLETE 8:25

-----  
Digital Signature #  
989E31B268BBA3DA  
-----