
Midmark M11 - v1.0.5
Total Cycles: 1160

Name: RITTER2
Sterilizer ID: V2313228
06/06/2022 01:20 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.5 F 0.0
2:00 168.4 F 3.0
4:00 185.1 F 7.2
6:00 203.5 F 12.2
8:00 220.6 F 6.8
10:00 232.7 F 12.1
12:00 235.8 F 9.7
14:00 236.8 F 9.1
16:00 249.6 F 15.5
18:00 261.6 F 22.5
20:00 272.8 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.1
1:00 273.0 F 29.8
2:00 271.8 F 29.1
3:00 271.8 F 29.6
4:00 273.1 F 29.8

Min 271.3 F 28.9
Max 273.1 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.7 F 0.0
5:00 229.5 F 0.0
10:00 255.3 F 0.0
15:00 232.8 F 0.0
20:00 228.2 F 0.0
25:00 227.2 F 0.0
30:00 225.1 F 0.0
Drying Complete

FILLING: 1:23
HEATING: 20:25
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:58
CYCLE COMPLETE

06/06/2022 02:18 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____