
Midmark M11 - v1.0.0
Total Cycles: 1699

Name: RITTER 1
Sterilizer ID: V2178198
03/09/2022 09:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.9 F 0.0
2:00 164.4 F 1.7
4:00 168.1 F 4.9
6:00 188.8 F 9.2
8:00 216.9 F 7.1
10:00 223.9 F 8.0
12:00 235.7 F 13.0
14:00 237.1 F 10.0
16:00 236.3 F 8.4
18:00 248.6 F 14.7
20:00 260.6 F 21.3
22:00 271.3 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 272.7 F 29.1
2:00 271.7 F 28.6
3:00 272.7 F 29.2
4:00 271.6 F 28.5

Min 271.6 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.0 F 0.0
5:00 228.8 F 0.0
10:00 226.8 F 0.0
15:00 229.7 F 0.0
20:00 232.2 F 0.0
25:00 227.3 F 0.0
30:00 228.2 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 22:39
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:00:47
CYCLE COMPLETE

03/09/2022 10:50 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____