

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3188

Name: RITTER 1  
Sterilizer ID: V2178198  
05/02/2023 12:22 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 178.0 F 0.0  
2:00 171.8 F 1.8  
4:00 190.7 F 5.9  
6:00 212.6 F 11.0  
8:00 221.4 F 6.3  
10:00 234.8 F 11.7  
12:00 235.1 F 9.2  
14:00 237.5 F 9.3  
16:00 248.4 F 14.8  
18:00 261.0 F 21.9  
20:00 272.5 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.7  
1:00 272.7 F 29.4  
2:00 271.8 F 28.6  
3:00 272.3 F 29.4  
4:00 272.5 F 29.3

Min 271.5 F 28.5  
Max 272.8 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.1 F 0.0  
5:00 218.0 F 0.0  
10:00 238.9 F 0.0  
15:00 236.5 F 0.0  
20:00 228.4 F 0.0  
25:00 227.2 F 0.0  
30:00 226.9 F 0.0  
Drying Complete

FILLING: 1:18  
HEATING: 20:27  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 30:00  
TOTAL CYCLE: 00:59:04  
CYCLE COMPLETE

05/02/2023 01:21 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_