

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4814

Name: RITTER 1  
Sterilizer ID: V2178198  
08/14/2024 09:24 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 163.5 F 0.0  
2:00 164.4 F 1.0  
4:00 176.6 F 2.0  
6:00 196.3 F 5.6  
8:00 210.7 F 9.2  
10:00 225.5 F 13.3  
12:00 225.9 F 7.7  
14:00 236.9 F 12.3  
16:00 235.3 F 9.2  
18:00 239.6 F 10.2  
20:00 246.0 F 13.2  
22:00 257.4 F 19.2  
24:00 267.4 F 25.3  
26:00 272.2 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.5  
1:00 271.8 F 28.7  
2:00 272.3 F 29.1  
3:00 272.7 F 29.3  
4:00 272.7 F 29.2

Min 271.4 F 28.3  
Max 272.8 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.1 F 0.0  
5:00 232.9 F 0.0  
10:00 230.8 F 0.0  
15:00 212.9 F 0.0  
20:00 198.6 F 0.0  
25:00 199.8 F 0.0  
30:00 201.2 F 0.0  
Drying Complete

FILLING: 1:17  
HEATING: 26:40  
STERILIZING: 4:00  
VENTING: 3:51  
DRYING: 30:00  
TOTAL CYCLE: 01:05:48  
CYCLE COMPLETE

08/14/2024 10:29 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_