
Midmark M11 - v1.0.0
Total Cycles: 845

Name: RITTER 1
Sterilizer ID: V2178198
05/27/2021 12:55 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 156.5 F 0.0
2:00 155.2 F 1.7
4:00 170.9 F 4.0
6:00 189.8 F 7.4
8:00 206.9 F 11.8
10:00 222.0 F 5.3
12:00 231.7 F 9.5
14:00 238.5 F 10.7
16:00 241.0 F 11.2
18:00 237.7 F 9.2
20:00 250.6 F 15.4
22:00 261.5 F 21.5
24:00 271.3 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.1
1:00 272.5 F 28.8
2:00 271.8 F 28.1
3:00 272.1 F 28.7
4:00 272.3 F 28.6

Min 271.3 F 28.0
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 215.2 F 0.0
5:00 217.9 F 0.0
10:00 210.7 F 0.0
15:00 205.2 F 0.0
20:00 205.8 F 0.0
25:00 219.6 F 0.0
30:00 217.6 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 24:42
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:03:18
CYCLE COMPLETE

05/27/2021 01:58 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____