
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.8uS / 0.5ppm
CYCLE NUMBER 001441
13:59 09/16/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:50
PRESSURIZING 5:31
STERILIZING 7:01
136.0 C 220kPa 7:01
136.4 C 222kPa 7:31
136.2 C 222kPa 8:01
136.4 C 222kPa 8:31
136.5 C 220kPa 9:01
136.3 C 222kPa 9:31
136.2 C 221kPa 10:01
136.4 C 223kPa 10:31
VENTING 10:31
AIR DRYING 11:36
CYCLE COMPLETE 32:27

Digital Signature #
EF0EA80ADD02FE1C
