
Midmark M11 - v1.0.5
Total Cycles: 561

Name: RITTER2
Sterilizer ID: V2313228
11/17/2021 04:02 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.6 F 0.0
2:00 161.8 F 3.0
4:00 174.2 F 7.1
6:00 199.3 F 12.0
8:00 216.2 F 6.0
10:00 228.9 F 10.8
12:00 230.6 F 7.6
14:00 241.1 F 12.3
16:00 243.7 F 13.0
18:00 255.9 F 19.6
20:00 266.8 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.6
1:00 272.7 F 30.2
2:00 271.6 F 29.5
3:00 272.0 F 30.4
4:00 272.2 F 30.0

Min 271.3 F 29.4
Max 272.9 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.4 F 0.0
5:00 197.7 F 0.0
10:00 202.4 F 0.0
15:00 202.6 F 0.0
20:00 201.4 F 0.0
25:00 198.9 F 0.0
30:00 197.4 F 0.0
Drying Complete

FILLING: 1:38
HEATING: 21:31
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:00:29
CYCLE COMPLETE

11/17/2021 05:02 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____