

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001339  
13:18 10/11/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:04  
PRESSURIZING 4:07  
STERILIZING 5:21  
136.0 C 221kPa 5:21  
136.4 C 222kPa 5:51  
136.4 C 225kPa 6:21  
136.1 C 222kPa 6:51  
136.4 C 225kPa 7:21  
136.4 C 223kPa 7:51  
136.3 C 224kPa 8:21  
136.3 C 223kPa 8:51  
136.3 C 223kPa 8:51  
VENTING 8:51  
AIR DRYING 9:44  
CYCLE COMPLETE 15:02  
-----

Digital Signature #  
4A206F384EF95BDB  
-----