
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000397
10:20 02/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:27
PRESSURIZING 2:34
STERILIZING 3:31
136.4 C 220kPa 3:31
136.3 C 221kPa 4:01
136.4 C 220kPa 4:31
136.2 C 224kPa 5:01
135.9 C 221kPa 5:31
136.3 C 222kPa 6:01
136.4 C 223kPa 6:31
136.4 C 223kPa 7:01
136.4 C 223kPa 7:01
VENTING 7:01
AIR DRYING 7:53
CYCLE COMPLETE 7:58

Digital Signature #
E3862C94807C6BDA
