
Midmark M11 - v1.0.5
Total Cycles: 2991

Name: RITTER2
Sterilizer ID: V2313228
10/10/2023 04:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 161.4 F 0.0
2:00 160.5 F 2.9
4:00 177.3 F 6.7
6:00 198.6 F 11.3
8:00 219.3 F 5.5
10:00 225.4 F 10.3
12:00 231.2 F 7.9
14:00 243.4 F 13.1
16:00 242.8 F 12.2
18:00 255.4 F 18.7
20:00 266.5 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.0
1:00 272.4 F 29.7
2:00 271.2 F 28.9
3:00 272.3 F 29.8
4:00 272.7 F 29.8

Min 271.0 F 28.9
Max 272.8 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.5 F 0.0
5:00 247.1 F 0.0
10:00 203.2 F 0.0
15:00 205.6 F 0.0
20:00 199.0 F 0.0
25:00 204.6 F 0.0
30:00 188.1 F 0.0
Drying Complete

FILLING: 1:18
HEATING: 21:38
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 01:00:22
CYCLE COMPLETE

10/10/2023 05:09 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____