
Midmark M11 - v1.0.0
Total Cycles: 4642

Name: RITTER 1
Sterilizer ID: V2178198
06/14/2024 09:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.7 F 0.0
2:00 187.9 F 2.3
4:00 184.9 F 5.8
6:00 201.8 F 10.0
8:00 221.9 F 6.8
10:00 225.3 F 7.9
12:00 235.1 F 12.2
14:00 235.2 F 8.8
16:00 241.0 F 11.1
18:00 245.2 F 12.7
20:00 256.2 F 18.7
22:00 266.7 F 25.0
24:00 272.8 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 271.8 F 28.6
2:00 272.1 F 28.6
3:00 272.6 F 28.9
4:00 272.4 F 29.1

Min 271.5 F 28.5
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.0 F 0.0
5:00 236.0 F 0.0
10:00 223.8 F 0.0
15:00 223.4 F 0.0
20:00 229.4 F 0.0
25:00 229.4 F 0.0
30:00 189.1 F 0.0
Drying Complete

FILLING: 1:25
HEATING: 24:18
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:03:31
CYCLE COMPLETE

06/14/2024 10:23 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____