

-----  
Mississauga, ON  
Canada  
-----

STATIM 5000 S501R751  
SN 511020J00004  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000805  
13:35 18/03/2022  
WRAPPED  
134 C for 6 min.  
CYCLE START 0:00  
CONDITIONING 3:42  
PRESSURIZING 12:40  
STERILIZING 14:00  
136.4 C 223kPa 14:00  
136.3 C 222kPa 14:30  
136.4 C 223kPa 15:00  
135.5 C 217kPa 15:30  
136.2 C 222kPa 16:00  
136.4 C 223kPa 16:30  
136.4 C 224kPa 17:00  
136.1 C 221kPa 17:30  
136.3 C 222kPa 18:00  
136.0 C 221kPa 18:30  
136.4 C 221kPa 19:00  
136.4 C 222kPa 19:30  
136.4 C 223kPa 20:00  
136.4 C 223kPa 20:00  
VENTING 20:00  
AIR DRYING 21:00  
CYCLE COMPLETE 50:38

-----  
Digital Signature #  
BB25B97B67FA9272  
-----