
Midmark M11 - v1.0.5
Total Cycles: 2455

Name: RITTER2
Sterilizer ID: V2313228
06/01/2023 03:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.1 F 0.0
2:00 172.0 F 3.7
4:00 184.4 F 8.1
6:00 208.1 F 11.5
8:00 226.0 F 8.9
10:00 234.9 F 10.1
12:00 240.7 F 12.2
14:00 240.5 F 11.0
16:00 254.0 F 17.9
18:00 265.6 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.2
1:00 272.3 F 29.5
2:00 272.2 F 29.8
3:00 271.8 F 29.3
4:00 272.4 F 30.0

Min 271.2 F 29.1
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.3 F 0.0
5:00 213.2 F 0.0
10:00 231.9 F 0.0
15:00 232.1 F 0.0
20:00 226.3 F 0.0
25:00 228.4 F 0.0
30:00 226.6 F 0.0
Drying Complete

FILLING: 1:35
HEATING: 19:39
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:30
CYCLE COMPLETE

06/01/2023 04:04 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____