
Midmark M11 - v1.0.5
Total Cycles: 2719

Name: RITTER2
Sterilizer ID: V2313228
08/01/2023 12:59 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.3 F 0.0
2:00 181.7 F 2.4
4:00 202.5 F 7.8
6:00 220.8 F 13.5
8:00 230.2 F 9.7
10:00 234.2 F 8.7
12:00 242.1 F 11.8
14:00 246.0 F 13.7
16:00 258.7 F 20.8
18:00 270.3 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.1
1:00 272.6 F 29.7
2:00 271.6 F 29.2
3:00 272.6 F 29.7
4:00 271.8 F 29.1

Min 271.6 F 29.0
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.7 F 0.0
5:00 209.4 F 0.0
10:00 218.5 F 0.0
15:00 221.2 F 0.0
20:00 221.1 F 0.0
25:00 226.3 F 0.0
30:00 221.5 F 0.0
Drying Complete

FILLING: 2:21
HEATING: 18:52
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:26
CYCLE COMPLETE

08/01/2023 01:57 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____