
Midmark M11 - v1.0.5
Total Cycles: 3924

Name: RITTER2
Sterilizer ID: V2313228
09/10/2024 01:42 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 119.9 F 0.0
2:00 120.6 F 0.1
4:00 125.3 F 0.7
6:00 161.8 F 3.8
8:00 185.9 F 7.4
10:00 198.3 F 11.5
12:00 217.8 F 4.8
14:00 228.1 F 9.1
16:00 239.3 F 13.6
18:00 238.5 F 10.8
20:00 237.0 F 8.9
22:00 248.9 F 15.0
24:00 259.6 F 21.0
26:00 269.6 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.4
1:00 272.6 F 29.5
2:00 272.6 F 30.0
3:00 271.7 F 29.3
4:00 272.8 F 29.8

Min 271.6 F 29.3
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.4 F 0.0
5:00 222.4 F 0.0
10:00 219.7 F 0.0
15:00 212.8 F 0.0
20:00 210.6 F 0.0
25:00 215.7 F 0.0
30:00 215.7 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 27:45
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:06:24
CYCLE COMPLETE

09/10/2024 02:48 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____