
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001249
17:43 06/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:41
PRESSURIZING 3:42
STERILIZING 4:50
136.0 C 220kPa 4:50
136.1 C 222kPa 5:20
136.3 C 224kPa 5:50
136.4 C 225kPa 6:20
136.1 C 221kPa 6:50
136.3 C 223kPa 7:20
136.1 C 222kPa 7:50
136.2 C 223kPa 8:20
136.2 C 223kPa 8:20
VENTING 8:20
AIR DRYING 9:15
CYCLE COMPLETE 9:24

Digital Signature #
76A7CA376350A948
