

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3712

Name: RITTER 1  
Sterilizer ID: V2178198  
08/31/2023 02:29 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 128.3 F 0.0  
2:00 129.1 F 1.2  
4:00 146.5 F 4.0  
6:00 167.7 F 7.5  
8:00 207.9 F 12.0  
10:00 219.0 F 5.1  
12:00 225.6 F 9.3  
14:00 235.6 F 10.9  
16:00 238.0 F 10.0  
18:00 236.5 F 8.6  
20:00 247.0 F 13.5  
22:00 258.2 F 19.6  
24:00 268.6 F 26.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.1  
1:00 272.2 F 28.4  
2:00 271.9 F 28.8  
3:00 271.5 F 28.1  
4:00 272.5 F 29.3

Min 271.0 F 28.0  
Max 272.6 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 188.5 F 0.0  
5:00 232.6 F 0.0  
10:00 218.2 F 0.0  
15:00 200.1 F 0.0  
20:00 190.5 F 0.0  
25:00 191.0 F 0.0  
30:00 235.7 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 25:11  
STERILIZING: 4:00  
VENTING: 3:25  
DRYING: 30:00  
TOTAL CYCLE: 01:03:29  
CYCLE COMPLETE

08/31/2023 03:32 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_