
Midmark M11 - v1.0.5
Total Cycles: 1398

Name: RITTER2
Sterilizer ID: V2313228
08/16/2022 11:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.6 F 0.0
2:00 172.8 F 0.3
4:00 191.2 F 4.0
6:00 211.9 F 9.1
8:00 230.2 F 9.2
10:00 237.8 F 10.7
12:00 234.6 F 8.5
14:00 240.6 F 10.5
16:00 248.3 F 14.5
18:00 260.4 F 21.4
20:00 271.3 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.8
1:00 272.2 F 29.2
2:00 271.6 F 29.2
3:00 271.9 F 29.0
4:00 271.6 F 29.3

Min 271.1 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.3 F 0.0
5:00 228.1 F 0.0
10:00 234.9 F 0.0
15:00 233.9 F 0.0
20:00 216.2 F 0.0
25:00 206.7 F 0.0
30:00 186.5 F 0.0
Drying Complete

FILLING: 1:48
HEATING: 20:39
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:40
CYCLE COMPLETE

08/16/2022 12:13 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____