
Midmark M11 - v1.0.0
Total Cycles: 4309

Name: RITTER 1
Sterilizer ID: V2178198
02/16/2024 11:38 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 142.6 F 0.0
2:00 141.9 F 1.7
4:00 154.5 F 4.5
6:00 179.6 F 7.5
8:00 199.3 F 11.7
10:00 218.0 F 3.7
12:00 221.9 F 7.3
14:00 231.3 F 11.2
16:00 229.8 F 6.9
18:00 241.8 F 11.9
20:00 238.8 F 9.8
22:00 250.5 F 15.4
24:00 260.6 F 21.1
26:00 270.2 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.2
1:00 272.5 F 29.0
2:00 272.4 F 28.7
3:00 272.6 F 29.3
4:00 271.9 F 28.6

Min 271.0 F 28.1
Max 273.0 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.5 F 0.0
5:00 222.4 F 0.0
10:00 235.3 F 0.0
15:00 224.9 F 0.0
20:00 225.9 F 0.0
25:00 228.1 F 0.0
30:00 211.1 F 0.0
Drying Complete

FILLING: 0:60
HEATING: 26:57
STERILIZING: 4:00
VENTING: 3:54
DRYING: 30:00
TOTAL CYCLE: 01:05:51
CYCLE COMPLETE

02/16/2024 12:43 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____