

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000934  
15:27 16/02/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:33  
PRESSURIZING 2:45  
STERILIZING 3:54  
136.0 C 227kPa 3:54  
136.1 C 233kPa 4:24  
136.4 C 232kPa 4:54  
136.4 C 233kPa 5:24  
136.3 C 232kPa 5:54  
136.0 C 229kPa 6:24  
136.4 C 231kPa 6:54  
136.4 C 232kPa 7:24  
136.4 C 232kPa 7:24  
VENTING 7:24  
AIR DRYING 8:16  
CYCLE COMPLETE 23:16  
-----

Digital Signature #  
66C15DB25B847525  
-----