
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 007588
10:57 21/06/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:21
PRESSURIZING 2:26
STERILIZING 3:33
136.4 C 221kPa 3:33
136.4 C 225kPa 4:03
136.4 C 225kPa 4:33
136.2 C 222kPa 5:03
136.1 C 223kPa 5:33
136.3 C 224kPa 6:03
136.3 C 223kPa 6:33
136.4 C 225kPa 7:03
136.4 C 225kPa 7:03
VENTING 7:03
AIR DRYING 7:54
CASSETTE REMOVED
CYCLE COMPLETE 16:04

Digital Signature #
56CE0037D76EA097
