

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1322

Name: RITTER 1  
Sterilizer ID: V2178198  
10/29/2021 12:23 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 141.8 F 0.0  
2:00 145.8 F 1.6  
4:00 170.6 F 4.6  
6:00 192.6 F 8.9  
8:00 210.6 F 8.3  
10:00 227.6 F 7.7  
12:00 238.7 F 12.5  
14:00 236.8 F 9.3  
16:00 238.2 F 9.3  
18:00 247.7 F 13.9  
20:00 259.8 F 20.7  
22:00 270.7 F 27.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.3  
1:00 272.4 F 28.6  
2:00 272.1 F 28.9  
3:00 272.1 F 28.3  
4:00 272.6 F 29.1

Min 271.4 F 28.1  
Max 272.8 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 219.1 F 0.0  
5:00 230.5 F 0.0  
10:00 203.5 F 0.0  
15:00 196.5 F 0.0  
20:00 189.3 F 0.0  
25:00 192.9 F 0.0  
30:00 189.7 F 0.0  
Drying Complete

FILLING: 1:23  
HEATING: 22:42  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 01:01:14  
CYCLE COMPLETE

10/29/2021 01:36 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02