

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2353

Name: RITTER 1  
Sterilizer ID: V2178198  
09/21/2022 04:09 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 153.5 F 0.0  
2:00 154.5 F 3.2  
4:00 172.3 F 7.1  
6:00 198.2 F 12.0  
8:00 215.6 F 5.6  
10:00 227.8 F 10.5  
12:00 230.4 F 7.1  
14:00 243.0 F 12.9  
16:00 241.5 F 11.4  
18:00 254.5 F 18.0  
20:00 266.1 F 25.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.7  
1:00 272.6 F 29.5  
2:00 271.8 F 28.8  
3:00 271.9 F 29.2  
4:00 272.5 F 29.4  
  
Min 271.4 F 28.6  
Max 272.6 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.8 F 0.0  
5:00 220.5 F 0.0  
10:00 216.7 F 0.0  
15:00 225.6 F 0.0  
20:00 234.8 F 0.0  
25:00 223.9 F 0.0  
30:00 225.5 F 0.0  
Drying Complete

FILLING: 1:03  
HEATING: 21:40  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:59:57  
CYCLE COMPLETE

09/21/2022 05:08 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_