
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000002
9:43 19/12/2019
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:32
PRESSURIZING 2:38
STERILIZING 3:30
136.4 C 220kPa 3:30
136.3 C 219kPa 4:00
136.3 C 223kPa 4:30
136.3 C 223kPa 5:00
136.1 C 221kPa 5:30
136.4 C 224kPa 6:00
136.4 C 224kPa 6:30
136.4 C 223kPa 7:00
136.4 C 223kPa 7:00
VENTING 7:00
AIR DRYING 7:47
CYCLE COMPLETE 34:28

Digital Signature #
542806A1EA282823
