
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 006123
14:41 15/08/2023
WRAPPED
135 C for 10 min.
CYCLE START 0:00
CONDITIONING 2:29
PRESSURIZING 4:42
STERILIZING 5:28
136.0 C 220kPa 5:28
135.9 C 220kPa 5:58
136.3 C 225kPa 6:28
136.4 C 225kPa 6:58
136.4 C 223kPa 7:28
136.4 C 224kPa 7:58
136.4 C 224kPa 8:28
136.4 C 225kPa 8:58
136.3 C 220kPa 9:28
136.4 C 225kPa 9:58
136.4 C 224kPa 10:28
136.4 C 223kPa 10:58
136.4 C 224kPa 11:28
136.4 C 225kPa 11:58
136.4 C 224kPa 12:28
136.1 C 219kPa 12:58
136.3 C 224kPa 13:28
136.3 C 224kPa 13:58
136.4 C 224kPa 14:28
136.4 C 221kPa 14:58
136.2 C 223kPa 15:28
136.2 C 223kPa 15:28
VENTING 15:29
AIR DRYING 16:22
CYCLE COMPLETE 16:32

Digital Signature #

A4E27020F8301A9C
