
Midmark M11 - v1.0.0
Total Cycles: 3553

Name: RITTER 1
Sterilizer ID: V2178198
07/20/2023 07:34 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.7 F 0.0
2:00 73.2 F 0.0
4:00 81.6 F 0.0
6:00 106.4 F 0.0
8:00 138.5 F 1.3
10:00 165.0 F 3.4
12:00 183.4 F 6.5
14:00 206.7 F 9.9
16:00 218.8 F 13.4
18:00 221.4 F 6.5
20:00 232.9 F 10.3
22:00 235.5 F 8.8
24:00 240.9 F 11.6
26:00 237.8 F 9.5
28:00 250.9 F 15.7
30:00 261.6 F 21.8
32:00 271.5 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 271.7 F 28.4
2:00 272.2 F 28.8
3:00 271.8 F 28.7
4:00 271.5 F 28.2

Min 271.4 F 28.1
Max 272.3 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.2 F 0.0
5:00 192.8 F 0.0
10:00 205.3 F 0.0
15:00 214.4 F 0.0
20:00 222.4 F 0.0
25:00 227.9 F 0.0
30:00 229.3 F 0.0

Drying Complete

FILLING: 0:46
HEATING: 32:33
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:10:31
CYCLE COMPLETE

07/20/2023 08:44 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____