

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3091

Name: RITTER 1  
Sterilizer ID: V2178198  
04/11/2023 12:38 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 171.9 F 0.0  
2:00 171.9 F 3.4  
4:00 186.2 F 7.5  
6:00 212.6 F 12.7  
8:00 223.2 F 7.3  
10:00 236.0 F 12.5  
12:00 236.2 F 9.6  
14:00 237.6 F 8.8  
16:00 248.1 F 14.5  
18:00 260.3 F 21.2  
20:00 270.9 F 28.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.7  
1:00 272.3 F 29.1  
2:00 272.0 F 29.1  
3:00 272.1 F 28.9  
4:00 272.2 F 29.2

Min 271.4 F 28.6  
Max 272.8 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.2 F 0.0  
5:00 208.1 F 0.0  
10:00 226.7 F 0.0  
15:00 232.3 F 0.0  
20:00 237.7 F 0.0  
25:00 230.6 F 0.0  
30:00 224.4 F 0.0  
Drying Complete

FILLING: 1:20  
HEATING: 20:43  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:59:17  
CYCLE COMPLETE

04/11/2023 01:37 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_