
Midmark M11 - v1.0.0
Total Cycles: 3072

Name: RITTER 1
Sterilizer ID: V2178198
04/04/2023 10:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 127.8 F 0.0
2:00 129.5 F 0.0
4:00 145.4 F 1.5
6:00 184.6 F 4.2
8:00 204.3 F 8.7
10:00 219.1 F 13.6
12:00 226.5 F 7.9
14:00 238.8 F 12.8
16:00 238.0 F 10.1
18:00 236.3 F 8.3
20:00 249.8 F 15.0
22:00 261.8 F 21.8
24:00 272.7 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.2
1:00 272.4 F 28.7
2:00 271.8 F 28.5
3:00 272.3 F 28.6
4:00 271.7 F 28.5

Min 271.5 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.3 F 0.0
5:00 216.1 F 0.0
10:00 236.3 F 0.0
15:00 234.2 F 0.0
20:00 227.6 F 0.0
25:00 225.8 F 0.0
30:00 225.8 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 24:22
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:02:19
CYCLE COMPLETE

04/04/2023 11:28 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____