
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000245
12:09 29/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:25
PRESSURIZING 3:28
STERILIZING 5:06
136.0 C 229kPa 5:06
136.4 C 233kPa 5:36
136.4 C 233kPa 6:06
136.3 C 233kPa 6:36
136.2 C 232kPa 7:06
136.1 C 233kPa 7:36
136.4 C 233kPa 8:06
136.3 C 232kPa 8:36
136.3 C 232kPa 8:36
VENTING 8:36
AIR DRYING 9:38
CYCLE COMPLETE 15:46

Digital Signature #
C331C3E868B96BF0
