

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 000065  
12:32 03/07/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:51  
PRESSURIZING 2:59  
STERILIZING 4:04  
136.0 C 222kPa 4:04  
136.2 C 224kPa 4:34  
136.4 C 226kPa 5:04  
136.3 C 225kPa 5:34  
136.2 C 225kPa 6:04  
136.2 C 225kPa 6:34  
136.2 C 224kPa 7:04  
136.2 C 223kPa 7:34  
136.2 C 223kPa 7:34  
VENTING 7:34  
AIR DRYING 8:27  
CYCLE COMPLETE 17:23  
-----

Digital Signature #  
90CB8CA82C51F64F  
-----