
Midmark M11 - v1.0.0
Total Cycles: 1161

Name: RITTER 1
Sterilizer ID: V2178198
09/13/2021 09:07 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.2 F 0.0
2:00 185.2 F 2.1
4:00 183.7 F 5.7
6:00 199.5 F 10.4
8:00 229.6 F 6.0
10:00 230.0 F 10.5
12:00 234.3 F 7.5
14:00 243.8 F 13.1
16:00 242.3 F 11.4
18:00 254.7 F 17.7
20:00 265.6 F 24.3
22:00 272.1 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.7
1:00 271.7 F 28.7
2:00 271.8 F 28.7
3:00 271.8 F 28.6
4:00 272.1 F 28.6

Min 271.6 F 28.6
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.8 F 0.0
5:00 240.1 F 0.0
10:00 232.2 F 0.0
15:00 227.3 F 0.0
20:00 228.6 F 0.0
25:00 230.4 F 0.0
30:00 222.1 F 0.0
Drying Complete

FILLING: 1:31
HEATING: 22:25
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:01:08
CYCLE COMPLETE

09/13/2021 10:08 AM
ACCEPTED: X
REJECTED:
Operator: U01