
Midmark M11 - v1.0.5
Total Cycles: 3236

Name: RITTER2
Sterilizer ID: V2313228
12/20/2023 11:18 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 158.5 F 0.0
2:00 154.1 F 1.1
4:00 175.3 F 4.6
6:00 197.9 F 8.7
8:00 214.3 F 13.4
10:00 229.9 F 7.8
12:00 234.5 F 12.9
14:00 239.0 F 10.3
16:00 236.0 F 8.5
18:00 250.0 F 15.3
20:00 261.3 F 21.7
22:00 271.5 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.0
1:00 272.0 F 28.9
2:00 272.5 F 29.2
3:00 272.5 F 29.5
4:00 272.1 F 29.6

Min 271.4 F 28.9
Max 273.0 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.7 F 0.0
5:00 196.6 F 0.0
10:00 197.2 F 0.0
15:00 194.3 F 0.0
20:00 192.1 F 0.0
25:00 188.1 F 0.0
30:00 185.7 F 0.0
Drying Complete

FILLING: 1:21
HEATING: 23:11
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 01:01:58
CYCLE COMPLETE

12/20/2023 12:19 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____