

Midmark M11 - v1.0.5
Total Cycles: 900

Name: RITTER2
Sterilizer ID: V2313228
03/24/2022 09:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.8 F 0.0
2:00 168.1 F 0.1
4:00 185.5 F 3.6
6:00 206.0 F 8.6
8:00 223.1 F 9.8
10:00 231.7 F 10.4
12:00 233.7 F 8.1
14:00 241.3 F 10.9
16:00 247.2 F 14.4
18:00 260.4 F 21.7
20:00 272.2 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.6 F 29.6
2:00 271.4 F 28.8
3:00 272.5 F 29.8
4:00 272.1 F 29.2

Min 271.2 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.5 F 0.0
5:00 215.2 F 0.0
10:00 220.6 F 0.0
15:00 225.3 F 0.0
20:00 225.9 F 0.0
25:00 228.3 F 0.0
30:00 198.1 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 20:30
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:58:45
CYCLE COMPLETE

03/24/2022 10:50 AM
ACCEPTED: X
REJECTED:
Operator: U02