
Midmark M11 - v1.0.0
Total Cycles: 1122

Name: RITTER 1
Sterilizer ID: V2178198
08/27/2021 07:05 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	71.2 F	0.0
2:00	75.9 F	0.0
4:00	81.7 F	0.0
6:00	111.4 F	0.0
8:00	152.4 F	0.1
10:00	182.3 F	0.1
12:00	199.6 F	0.2
14:00	212.0 F	2.5
16:00	220.1 F	5.2
18:00	227.2 F	7.7
20:00	235.8 F	11.2
22:00	233.3 F	7.4
24:00	243.6 F	12.6
26:00	241.2 F	10.7
28:00	238.2 F	9.3
30:00	251.8 F	15.8
32:00	262.8 F	22.1
34:00	272.7 F	28.6

STERILIZING

mm:ss	Degrees	PSI
0:00	271.5 F	27.9
1:00	272.7 F	28.6
2:00	271.7 F	27.8
3:00	272.2 F	28.5
4:00	272.5 F	28.4

Min 271.5 F 27.7
Max 272.7 F 28.7

VENTING CHAMBER

DRYING

Drying Start	mm:ss	Degrees	PSI
	0:00	208.4 F	0.0
	5:00	220.7 F	0.0
	10:00	220.5 F	0.0
	15:00	211.5 F	0.0
	20:00	212.8 F	0.0
	25:00	219.8 F	0.0
Drying Complete	30:00	225.3 F	0.0

FILLING: 0:60
HEATING: 34:23
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:12:44
CYCLE COMPLETE

08/27/2021 08:17 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____