
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 005018
16:18 30/01/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:34
PRESSURIZING 2:36
STERILIZING 3:33
136.0 C 220kPa 3:33
136.2 C 218kPa 4:03
136.3 C 222kPa 4:33
136.1 C 222kPa 5:03
136.4 C 224kPa 5:33
136.3 C 224kPa 6:03
136.4 C 224kPa 6:33
136.3 C 221kPa 7:03
136.3 C 221kPa 7:03
VENTING 7:03
AIR DRYING 8:02
CYCLE COMPLETE 10:35

Digital Signature #
B55A8DF399B64AB2
