

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1425

Name: RITTER 1  
Sterilizer ID: V2178198  
12/06/2021 12:47 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 176.2 F 0.0  
2:00 175.4 F 2.4  
4:00 199.4 F 7.2  
6:00 216.5 F 12.4  
8:00 226.6 F 7.3  
10:00 238.7 F 12.6  
12:00 238.8 F 10.2  
14:00 235.1 F 8.3  
16:00 250.5 F 15.7  
18:00 263.1 F 22.9  
20:00 271.7 F 28.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.5  
1:00 272.3 F 28.9  
2:00 272.2 F 29.1  
3:00 271.9 F 28.7  
4:00 272.4 F 29.2  
  
Min 271.4 F 28.3  
Max 272.8 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 179.1 F 0.0  
5:00 238.7 F 0.0  
10:00 225.9 F 0.0  
15:00 208.4 F 0.0  
20:00 195.0 F 0.0  
25:00 198.4 F 0.0  
30:00 208.0 F 0.0  
Drying Complete

FILLING: 2:14  
HEATING: 20:04  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:59:28  
CYCLE COMPLETE

12/06/2021 02:15 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02