
Midmark M11 - v1.0.5
Total Cycles: 3697

Name: RITTER2
Sterilizer ID: V2313228
06/13/2024 03:57 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 124.4 F 0.0
2:00 128.3 F 1.8
4:00 148.4 F 4.7
6:00 175.1 F 8.3
8:00 194.0 F 12.5
10:00 212.3 F 5.0
12:00 224.7 F 9.3
14:00 233.0 F 11.3
16:00 237.7 F 11.0
18:00 236.0 F 9.2
20:00 249.1 F 15.4
22:00 260.3 F 21.7
24:00 270.6 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 29.1
1:00 272.6 F 29.9
2:00 271.8 F 29.1
3:00 272.5 F 29.9
4:00 272.0 F 29.3

Min 271.5 F 29.0
Max 272.9 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.0 F 0.0
5:00 208.6 F 0.0
10:00 198.7 F 0.0
15:00 195.6 F 0.0
20:00 193.0 F 0.0
25:00 191.3 F 0.0
30:00 195.7 F 0.0
Drying Complete

FILLING: 1:00
HEATING: 24:41
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:03:13
CYCLE COMPLETE

06/13/2024 05:00 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____