
Midmark M11 - v1.0.5
Total Cycles: 1968

Name: RITTER2
Sterilizer ID: V2313228
01/24/2023 04:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.8 F 0.0
2:00 180.4 F 3.3
4:00 200.2 F 8.3
6:00 217.1 F 11.0
8:00 228.6 F 9.2
10:00 234.7 F 9.6
12:00 244.5 F 13.4
14:00 244.3 F 12.8
16:00 257.5 F 19.9
18:00 269.2 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.2 F 29.3
2:00 271.5 F 29.1
3:00 272.3 F 29.4
4:00 271.4 F 28.8

Min 271.4 F 28.7
Max 272.5 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.9 F 0.0
5:00 237.9 F 0.0
10:00 211.2 F 0.0
15:00 209.0 F 0.0
20:00 210.9 F 0.0
25:00 213.0 F 0.0
30:00 217.1 F 0.0
Drying Complete

FILLING: 1:48
HEATING: 19:00
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:04
CYCLE COMPLETE

01/24/2023 04:59 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____