
Midmark M11 - v1.0.5
Total Cycles: 3504

Name: RITTER2
Sterilizer ID: V2313228
04/08/2024 11:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 130.6 F 0.0
2:00 132.0 F 1.2
4:00 139.2 F 3.7
6:00 166.1 F 7.1
8:00 189.2 F 11.3
10:00 208.6 F 4.1
12:00 219.1 F 8.5
14:00 232.0 F 13.3
16:00 234.6 F 9.9
18:00 236.3 F 9.2
20:00 246.0 F 14.0
22:00 257.8 F 20.2
24:00 268.2 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.3
1:00 272.4 F 30.1
2:00 272.1 F 29.6
3:00 271.0 F 29.2
4:00 272.1 F 30.0

Min 271.0 F 29.1
Max 272.7 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.7 F 0.0
5:00 213.4 F 0.0
10:00 207.7 F 0.0
15:00 196.4 F 0.0
20:00 214.5 F 0.0
25:00 220.1 F 0.0
30:00 223.8 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 25:22
STERILIZING: 4:00
VENTING: 3:28
DRYING: 30:00
TOTAL CYCLE: 01:03:47
CYCLE COMPLETE

04/08/2024 01:02 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____