

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.8uS / 0.5ppm  
CYCLE NUMBER 000600  
8:21 26/08/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:04  
PRESSURIZING 3:13  
STERILIZING 4:32  
136.0 C 227kPa 4:32  
136.2 C 231kPa 5:02  
136.4 C 230kPa 5:32  
136.3 C 229kPa 6:02  
136.4 C 231kPa 6:32  
136.4 C 231kPa 7:02  
136.3 C 231kPa 7:32  
136.4 C 232kPa 8:02  
136.4 C 232kPa 8:02  
VENTING 8:02  
AIR DRYING 8:57  
CYCLE COMPLETE 23:57  
-----

Digital Signature #  
B476BE6F6040991A  
-----