
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000872
13:28 17/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:08
PRESSURIZING 3:18
STERILIZING 4:42
136.0 C 222kPa 4:42
136.2 C 230kPa 5:12
136.1 C 230kPa 5:42
136.2 C 228kPa 6:12
136.2 C 231kPa 6:42
136.3 C 231kPa 7:12
136.3 C 231kPa 7:42
136.3 C 231kPa 8:12
136.3 C 231kPa 8:12
VENTING 8:12
AIR DRYING 9:07
CYCLE COMPLETE 22:35

Digital Signature #
A3456804AB676E9C
