
Midmark M11 - v1.0.0
Total Cycles: 1135

Name: RITTER 1
Sterilizer ID: V2178198
08/31/2021 10:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 150.1 F 0.0
2:00 148.4 F 1.1
4:00 161.8 F 3.1
6:00 185.3 F 5.8
8:00 202.3 F 9.6
10:00 218.8 F 11.2
12:00 222.4 F 6.8
14:00 232.8 F 11.2
16:00 234.2 F 7.7
18:00 244.0 F 13.4
20:00 243.0 F 12.0
22:00 254.6 F 18.0
24:00 265.0 F 24.2
26:00 271.3 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 272.1 F 29.2
2:00 272.7 F 29.3
3:00 272.2 F 29.0
4:00 271.6 F 28.7

Min 271.2 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.9 F 0.0
5:00 201.2 F 0.0
10:00 196.3 F 0.0
15:00 185.0 F 0.0
20:00 183.8 F 0.0
25:00 187.4 F 0.0
30:00 198.3 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 26:04
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:04:50
CYCLE COMPLETE

08/31/2021 11:36 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____