

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4081

Name: RITTER 1  
Sterilizer ID: V2178198  
12/04/2023 05:31 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 130.8 F 0.0  
2:00 140.4 F 1.9  
4:00 155.5 F 5.3  
6:00 179.6 F 9.5  
8:00 203.7 F 7.1  
10:00 215.9 F 6.4  
12:00 228.4 F 10.9  
14:00 229.9 F 6.7  
16:00 241.9 F 12.4  
18:00 240.2 F 10.6  
20:00 253.1 F 17.0  
22:00 264.4 F 23.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.4  
1:00 272.5 F 29.0  
2:00 271.5 F 28.4  
3:00 272.5 F 29.1  
4:00 271.8 F 28.4

Min 271.4 F 28.3  
Max 272.6 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 181.7 F 0.0  
5:00 212.3 F 0.0  
10:00 216.1 F 0.0  
15:00 217.4 F 0.0  
20:00 215.6 F 0.0  
25:00 217.1 F 0.0  
30:00 220.2 F 0.0  
Drying Complete

FILLING: 1:11  
HEATING: 23:57  
STERILIZING: 4:00  
VENTING: 3:22  
DRYING: 30:00  
TOTAL CYCLE: 01:02:31  
CYCLE COMPLETE

12/04/2023 06:33 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_