
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 005066
10:43 06/02/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:56
PRESSURIZING 1:59
STERILIZING 2:51
136.0 C 220kPa 2:51
136.0 C 220kPa 3:21
136.1 C 222kPa 3:51
136.4 C 223kPa 4:21
136.3 C 222kPa 4:51
136.4 C 224kPa 5:21
136.4 C 224kPa 5:51
136.4 C 223kPa 6:21
136.4 C 223kPa 6:21
VENTING 6:21
AIR DRYING 7:12
CYCLE COMPLETE 9:35

Digital Signature #
A3EFD86E2D3962D7
