
Midmark M11 - v1.0.5
Total Cycles: 461

Name: RITTER2
Sterilizer ID: V2313228
10/19/2021 11:17 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 203.0 F 0.0
2:00 199.8 F 4.3
4:00 210.3 F 10.4
6:00 224.1 F 7.1
8:00 239.0 F 13.4
10:00 241.5 F 12.3
12:00 242.6 F 12.1
14:00 256.5 F 19.4
16:00 268.5 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.1
1:00 272.5 F 29.4
2:00 272.3 F 29.7
3:00 272.2 F 29.2
4:00 272.5 F 29.9

Min 271.6 F 28.9
Max 272.9 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.9 F 0.0
5:00 191.0 F 0.0
10:00 197.3 F 0.0
15:00 192.5 F 0.0
20:00 187.8 F 0.0
25:00 186.6 F 0.0
30:00 183.3 F 0.0
Drying Complete

FILLING: 2:44
HEATING: 17:06
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:56:56
CYCLE COMPLETE

10/19/2021 12:42 PM
ACCEPTED: X
REJECTED:
Operator: U01