

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2549

Name: RITTER 1  
Sterilizer ID: V2178198  
11/14/2022 09:36 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 73.9 F 0.0  
2:00 76.8 F 0.0  
4:00 103.2 F 0.0  
6:00 135.6 F 1.5  
8:00 161.0 F 4.2  
10:00 185.4 F 7.6  
12:00 204.7 F 11.2  
14:00 213.5 F 3.6  
16:00 219.4 F 6.6  
18:00 229.6 F 10.2  
20:00 236.2 F 8.9  
22:00 237.5 F 10.6  
24:00 237.9 F 7.8  
26:00 247.3 F 13.9  
28:00 258.4 F 19.9  
30:00 268.5 F 26.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.4  
1:00 272.0 F 28.9  
2:00 272.3 F 28.8  
3:00 271.5 F 28.6  
4:00 272.5 F 29.1

Min 271.4 F 28.3  
Max 272.6 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.8 F 0.0  
5:00 212.4 F 0.0  
10:00 222.7 F 0.0  
15:00 228.1 F 0.0  
20:00 231.2 F 0.0  
25:00 230.4 F 0.0  
30:00 227.7 F 0.0  
Drying Complete

FILLING: 0:48  
HEATING: 31:07  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:09:07  
CYCLE COMPLETE

11/14/2022 10:45 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_