
Midmark M11 - v1.0.5
Total Cycles: 720

Name: RITTER2
Sterilizer ID: V2313228
01/24/2022 08:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.5 F 0.0
2:00 180.6 F 2.5
4:00 198.0 F 7.7
6:00 218.8 F 12.3
8:00 229.5 F 9.3
10:00 234.8 F 9.2
12:00 243.6 F 12.3
14:00 246.7 F 13.8
16:00 260.2 F 21.3
18:00 272.0 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.7
1:00 272.7 F 29.2
2:00 271.7 F 29.0
3:00 272.5 F 29.2
4:00 271.5 F 28.8

Min 271.5 F 28.5
Max 272.9 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.5 F 0.0
5:00 218.1 F 0.0
10:00 206.1 F 0.0
15:00 206.8 F 0.0
20:00 205.2 F 0.0
25:00 202.8 F 0.0
30:00 200.3 F 0.0
Drying Complete

FILLING: 1:54
HEATING: 18:30
STERILIZING: 4:00
VENTING: 3:02
DRYING: 30:00
TOTAL CYCLE: 00:57:25
CYCLE COMPLETE

01/24/2022 09:46 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____