
Midmark M11 - v1.0.5
Total Cycles: 1059

Name: RITTER2
Sterilizer ID: V2313228
05/03/2022 06:53 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.6 F 0.0
2:00 72.4 F 0.0
4:00 81.5 F 0.0
6:00 120.0 F 0.0
8:00 143.6 F 0.2
10:00 173.1 F 1.4
12:00 190.1 F 4.9
14:00 208.4 F 8.2
16:00 220.0 F 11.8
18:00 225.3 F 6.2
20:00 236.9 F 10.7
22:00 234.8 F 8.3
24:00 244.8 F 13.4
26:00 243.4 F 12.2
28:00 255.4 F 18.4
30:00 265.8 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.7 F 29.4
2:00 271.5 F 28.6
3:00 272.6 F 29.5
4:00 272.1 F 28.8

Min 271.3 F 28.6
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.8 F 0.0
5:00 207.9 F 0.0
10:00 186.6 F 0.0
15:00 191.3 F 0.0
20:00 196.6 F 0.0
25:00 184.7 F 0.0
Dry Cycle Aborted

FILLING: 0:47
HEATING: 31:48
STERILIZING: 4:00
VENTING: 3:07
DRYING: 29:30
TOTAL CYCLE: 01:09:12

05/03/2022 08:02 AM
ACCEPTED: X
REJECTED:
Operator: U02