
Midmark M11 - v1.0.0
Total Cycles: 4029

Name: RITTER 1
Sterilizer ID: V2178198
11/16/2023 11:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.3 F 0.0
2:00 176.0 F 2.4
4:00 190.5 F 5.7
6:00 196.4 F 10.5
8:00 217.8 F 5.2
10:00 220.5 F 8.4
12:00 231.9 F 12.9
14:00 234.8 F 9.5
16:00 238.5 F 9.7
18:00 246.4 F 13.4
20:00 257.7 F 19.4
22:00 268.0 F 25.8
24:00 272.6 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.8 F 29.2
2:00 271.5 F 28.5
3:00 272.7 F 29.3
4:00 271.9 F 28.7

Min 271.5 F 28.5
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.4 F 0.0
5:00 213.5 F 0.0
10:00 225.0 F 0.0
15:00 223.2 F 0.0
20:00 218.9 F 0.0
25:00 221.9 F 0.0
30:00 222.9 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 24:17
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:03:16
CYCLE COMPLETE

11/16/2023 12:35 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____